


Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 Accredited to ISO/IEC 17025:2005	Clearwater Environmental Ltd Issue No: 002 Issue date: 17 March 2008	
	Clearwater House River Way Harlow Essex CM20 2GE	Contact: Mr D Marston Tel: +44 (0)1279 423370 Fax: +44 (0)1279 423401 E-Mail: lab@clearwater-environmental.com Website: www.clearwater-environmental.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Clearwater House River Way Harlow Essex CM20 2GE Local contact Mr D Marston Tel: +44 (0)1279 423370 Fax: +44 (0)1279 423401 Email: lab@clearwater-environmental.com	Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises, including domestic, commercial and industrial	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



4129
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Clearwater Environmental Ltd
Issue No: 002 Issue date: 17 March 2008

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	A, B, C
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	A, B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:2005 by stereo-microscopy, polarised light microscopy and dispersion staining, and In-House Method PR02	A
	Sampling of bulk materials for asbestos identification	Documented In-House Method PR01	B
END			